

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130122 ✓
Name	Kirsten Hansen	ID#	5440 ✓
Date	09/16/2010 ✓		
A Agency	DJT0/ Alaska State Troopers	Phone #	(907) 895-4800
Instrument Location	Delta Jct. Court House		
B Alco S/N	75602 ✓	Target Value	.081 ✓
		High Pressure	915 psi
Alco Test Values	0.080 ✓	0.079 ✓	
	1 st Alco	2 nd Alco	
Signature	<u>Kirsten Hansen</u>		

(OVER)

(Do Not write in the area below)

SMB
12/9/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15th day of DEC, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
Carolyn M. Noland DANIEL L. NICHOLAS
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130122

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

SEPTEMBER 16, 2010

OPERATOR'S NAME:
HANGEN/KIRSTEN/Z
OPERATOR'S NUMBER: 5440
SUBJECT'S LAST NAME:
TEST

SUBJECT'S FIRST NAME/MI :
SEPT

O.L. #: 20100916
DEPT/AGENCY: DJT0
CASE/REPORT: 1000009
TEST TYPE: U
ALCO TARGET VALUE: .001
ALCO S/N: 75602

— BREATH ANALYSIS —

.001 ADJUSTED FOR 29.15 in
ALCO TARGET .078 17:21
BLANK TEST .000 17:22
INTERNAL STANDARD VERIFIED 17:22
ALCO TV 29.15 in .000 17:22
BLANK TEST .000 17:23
SUBJECT SAMPLE .000 17:24
BLANK TEST .000 17:24
ALCO TV 29.13 in .079 17:25
BLANK TEST .000 17:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

SEPTEMBER 16, 2010
TIME 17:17

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.15 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*~./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
~!@#\$%^&*~./0123456789:;<=>?@ABCDEFG